
OIV-MA-AS312-03A Methanol

Type II method

1. Scope of application

This method is applicable to the determination of methanol in wine for concentrations between 50 and 500 mg/L.

2. Principle

Methanol is determined in the distillate, to which an internal standard is added, using gas chromatography with a flame ionisation detector (FID).

3. Reagents and materials

3.1. Type II water, according to ISO standard 3696

3.2. Ethanol: purity \geq 96 % (CAS no. 64-17-5)

3.3. Hydrogen: minimum specifications: 99.999% purity (CAS no. 1333-74-0)

3.4. Helium: minimum specifications: 99.999% purity (CAS no. 7440-59-7)

3.5. Methanol: purity \geq 99 % (CAS no. 67-56-1)

3.6. 4-Methyl-2-pentanol (internal standard): purity \geq 98 % (CAS no. 108-11-2).
Internal standard used in the validation.

Note 1: Other internal standards can be used, such as:

3-pentanol: purity \geq 98% (CAS no. 584-02-1)

4-methyl-1-pentanol: purity \geq 98% (CAS no. 626-89-1)

Methyl nonanoate: purity \geq 98% (CAS no. 1731-84-6)

3.7. Reference materials: these may be, for example, wines from laboratory proficiency tests.

3.8. Preparation of working solutions (by way of example):

3.8.1. *Approximately 10% v/v aqueous-alcoholic mixture*

This mixture should be as close as possible to the alcohol content of the wine to be analysed. Pour 100 mL of ethanol (3.2) into a 1 L calibrated flask (4.2), make up to volume with demineralised water (3.1) and mix.

3.8.2. *10 g/L Internal standard solution*

Using an analytical balance (4.1), weigh approximately 1 g of internal standard (3.6) into a 100 mL calibrated flask (4.3) that contains around 60 mL of 10% ethanol solution

Methanol (GC) (Type-IV)

(3.8.1), so as to minimise evaporation of the internal standard. Make up to volume with the ethanol solution (3.8.1) and mix.

3.8.3. *1 g/L Internal standard solution*

Add 10 mL of the 10 g/L internal standard solution (3.8.2) using a pipette (4.8) and make up to 100 mL (4.3) using the 10% v/v hydroalcoholic mixture (3.8.1).

3.8.4. *5 g/L Methanol stock solution*

Using an analytical balance (4.1), weigh approximately 500 mg of methanol (3.5) into a 100 mL calibrated flask (4.3) that contains about 60 mL of 10% ethanol solution (3.8.1), so as to minimise evaporation of the methanol. Make up to volume with the ethanol solution (3.8.1) and mix.

3.8.5. *Working calibration solutions*

By way of example, a method for plotting a calibration curve is outlined below.

Each solution should be prepared with the 10% aqueous-alcoholic mixture (3.8.1).

3.8.5.1. *500 mg/L Methanol standard solution*

Add 10 mL of the 5 g/L stock solution (3.8.4) to a 100 mL calibrated flask (4.3) using a pipette (4.8) and make up to volume with the 10% v/v ethanol solution (3.8.1).

3.8.5.1.1. *250 mg/L Methanol standard solution*

Add 10 mL of the 500 mg/L methanol solution (3.8.5.1) to a 20 mL calibrated flask (4.5) using a pipette (4.8) and make up to volume with the 10% v/v ethanol solution (3.8.1).

3.8.5.1.2. *200 mg/L Methanol standard solution*

Add 20 mL of the 500 mg/L methanol solution (3.8.5.1) to a 50 mL calibrated flask (4.4) using a pipette (4.7) and make up to volume with the 10% v/v ethanol solution (3.8.1).

3.8.5.1.3. *150 mg/L Methanol standard solution*

Add 6 mL of the 500 mg/L methanol solution (3.8.5.1) to a 20 mL calibrated flask (4.5) using a pipette (4.9) and make up to volume with the 10% v/v ethanol solution (3.8.1).

3.8.5.1.4. *100 mg/L Methanol standard solution*

Add 4 mL of the 500 mg/L methanol solution (3.8.5.1) to a 20 mL calibrated flask (4.5) using a pipette (4.10) and make up to volume with the 10% v/v ethanol solution (3.8.1).

3.8.5.1.5. *50 mg/L Methanol standard solution*

Add 2 mL of the 500 mg/L methanol solution (3.8.5.1) to a 20 mL calibrated flask (4.5) using a pipette (4.11) and make up to volume with the 10% v/v ethanol solution (3.8.1).

4. Apparatus

4.1. Analytical balance (1 mg precision)

- 4.2. 1 L Class A calibrated flasks
- 4.3. 100 mL Class A calibrated flasks
- 4.4. 50 mL Class A calibrated flasks
- 4.5. 20 mL Class A calibrated flasks
- 4.6. 10 mL Class A calibrated flasks
- 4.7. 20 mL Class A pipettes with two marks
- 4.8. 10 mL Class A pipettes with two marks
- 4.9. 6 mL Class A pipettes with two marks
- 4.10. 4 mL Class A pipettes with two marks
- 4.11. 2 mL Class A pipettes with two marks
- 4.12. 1 mL Class A pipettes with two marks or 1 mL micropipettes
- 4.13. Temperature-programmable gas chromatograph with a flame ionisation detector and a data processing system capable of calculating areas or measuring peak heights
- 4.14. Fused silica capillary column coated with a Carbowax 20M-type polar stationary phase (for example):
 - Chrompack CP-wax 57 CB, 50 m x 0.32 mm x 0.45 μm
 - DB-WAX 52, 30 m x 25 μm x 0.2 μm

5. Sample preparation

Sparkling and/or young wines must be pre-degassed, for example, by mixing 200 mL of wine in a 1 L flask. Subsequently, the samples are distilled according to the method for determining alcoholic strength by volume (OIV-MA-AS312-01). The distillation can be carried out without adding calcium hydroxide in this case.

5.1. Addition of internal standard (by way of example)

Pour 10 mL of distillate into a 10 mL calibrated flask (4.6), add 1 mL (4.12) of internal standard solution (3.8.3) and mix.

6. Procedure

The calibration curve standards are treated in the same way as the samples (point 5.1). It is recommended that the aqueous-alcoholic mixture (3.8.1) is injected at the start of

Methanol (GC) (Type-IV)

the sequence in order to verify that it does not contain methanol.

6.1. Operating conditions (as a guide):

Carrier gas: helium or hydrogen

Carrier gas flow: 7 mL/min

Injection: split (ratio: 7:50)

Injection volume: 1 or 2 µL

Injector temperature: 200-260 °C

Detector temperature: 220-300 °C

Temperature programme: from 35 °C (for 2 minutes) to 170 °C, at 7.5 °C/min

7. Calculations

Calculate the concentration of methanol (C_i), using the following equation:

$$C_i = \frac{C_p}{m} \left(\frac{A_i}{A_p} - b \right)$$

A_i - Peak area of methanol

A_p - Peak area of internal standard

C_p - Concentration of internal standard

m - Slope of the calibration curve

b - Y-intercept of the calibration curve

8. Expression of the results

The concentration of methanol may be expressed in mg/L or in mg/100 mL absolute alcohol; in the latter case, the alcohol content by volume of the wine should be determined.

Note 2: mg/100 mL absolute alcohol = mg/L x 10/alcohol content by volume

9. Precision

The data from the international interlaboratory test is outlined in Annex A.

10. Quality control

Internal quality control may be carried out using certified reference materials or wines whose characteristics have been determined from a consensus (3.7). These should be prepared as for the samples (point 5). Participation in proficiency tests is

COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS

Methanol (GC) (Type-IV)

recommended.

11. Report of the results

The results are expressed to the nearest whole number (in accordance with the uncertainty).

12. Bibliography

Compendium of international methods of wine and must analysis. Method OIV-MA-AS312-01 (Alcoholic strength).

Annex A Statistical results of the interlaboratory test

Design of validation study

The validation study was conducted with 10 samples: 2 white wines, one dry and one sweet, 2 red wines, one of which was oaked, and 1 fortified wine (Port), including blind duplicates, according to OIV recommendations. The approximate concentration of methanol is shown in the following table.

| Sample | White wine Dry | White wine Sweet | Red wine | Red wine oaked | Fortified wine port |
|--------------------|-------------------|---------------------|----------|-------------------|---------------------------|
| Methanol (mg/L) | 50 | 150* | 270 | 400* | 120 |

(*) In this particular indicated case, methanol was added to the wine to cover a greater range of concentrations. The wine was then mixed, stabilised and bottled.

Participating laboratories:

Samples were sent to 17 laboratories in 9 different countries.

Laboratorios Agroalimentarios, Madrid (Spain)

Estación de Viticultura y Enología de Galicia, EVEGA (Spain)

Estació de Viticultura i Enologia de Vilafranca del Penedès, (Spain)

Estación Enológica de Haro, La Rioja (Spain)

Estación de Viticultura y Enología de Galicia (Spain)

Lab. Bordeaux, Service Commun des Lab., Pessac (France)

COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS

Methanol (GC) (Type-IV)

Laboratoire d'Ile-de-France, Paris (France)

Laboratoires Inter Rhône (France)

Comité Interprof. du Vin de Champagne (CIVC) (France)

Bfr-Bundesinst. f. Risikobewertung (Germany)

Landesuntersuchungsamt Mainz (Germany)

Instituto Nacional de Vitivinicultura, Mendoza (Argentina)

ALKO Inc., Alcohol Control Lab. (ACL) (Finland)

Instituto dos Vinhos do Douro e do Porto (Portugal)

Collaborative study on methanol

| Laboratory code | Dry white | | Sweet white | | Red | | Oaked red | | Port | |
|-----------------|--------------|--------------|---------------|---------------|--------|--------|---------------|---------------|--------|--------|
| | A | G | B | H | C | I | D | J | E | K |
| A | 39.99 | 38.13 | 127.42 | 136.25 | 144.80 | 145.71 | 496.53 | 513.00 | 192.13 | 219.39 |
| B | 41.20 | 40.90 | 157.60 | 160.50 | 150.40 | 146.90 | 484.90 | 477.80 | 222.40 | 219.60 |
| C | 36.80 | 35.60 | 133.50 | 129.20 | 119.10 | 134.10 | 454.10 | 478.40 | 197.00 | 174.80 |
| D | 36.00 | 39.60 | 177.40 | 145.50 | 160.80 | 138.00 | 302.00 | 494.50 | 216.10 | 248.50 |
| E | 68.00 | 70.00 | 163.00 | 169.00 | 178.00 | 177.00 | 503.00 | 495.50 | 225.00 | 227.00 |
| F | 37.00 | 37.10 | 148.30 | 148.20 | 143.40 | 142.40 | 484.10 | 474.00 | 206.30 | 206.90 |
| G | 41.40 | 42.30 | 152.60 | 152.40 | 149.70 | 150.50 | 489.60 | 491.10 | 216.60 | 217.20 |
| H | 36.80 | 32.40 | 140.80 | 129.10 | 128.00 | 137.70 | 440.60 | 429.30 | 187.50 | 192.80 |
| I | 42.90 | 43.30 | 153.50 | 155.50 | 139.70 | 147.40 | 468.30 | 456.10 | 225.30 | 225.60 |
| J | 40.90 | 40.60 | 155.50 | 154.60 | 148.50 | 149.10 | 496.40 | 499.80 | 217.10 | 217.00 |
| K | 39.30 | 36.20 | 103.10 | 143.10 | 131.90 | 115.90 | 437.90 | 334.00 | 156.10 | 172.60 |
| L | 35.00 | 39.00 | 164.00 | 167.00 | 157.00 | 160.00 | 492.00 | 508.00 | 249.00 | 220.00 |
| M | 43.60 | 43.40 | 157.30 | 154.90 | 155.50 | 158.90 | 506.80 | 496.10 | 217.70 | 219.50 |
| N | 34.20 | 33.60 | 126.50 | 125.70 | 125.90 | 133.60 | 429.10 | 429.00 | 192.10 | 188.90 |
| O | 34.00 | 35.70 | 149.00 | 154.80 | 144.20 | 141.80 | 482.80 | 473.60 | 210.40 | 218.10 |
| P | 44.70 | 43.70 | 151.60 | 146.90 | 140.70 | 147.60 | 451.20 | 472.80 | 205.40 | 205.80 |
| Q | 40.70 | 38.80 | 153.00 | 149.80 | 158.20 | 153.40 | 498.20 | 497.50 | 225.50 | 217.20 |

Czech Agriculture and Food Inspection Authority (CAFIA), Brno (Czech Republic) CZ
 National Food Safety Office, Directorate of Oenology and Alcoholic Beverages (NÉBIH BAI), Budapest (Hungary)
 Lehr- und Forschungszentrum, Klosterneuburg (Austria)

COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS

Methanol (GC) (Type-IV)

| Indicators | Dry white | Sweet white | Red | Oaked red | Port |
|------------------------------------|-----------|-------------|-----------|-----------|-----------|
| Number of accepted laboratories | 16 | 15 | 17 | 15 | 17 |
| Number of repetitions | 2 | 2 | 2 | 2 | 2 |
| Minimum | 32.40 | 125.70 | 115.90 | 429.00 | 156.10 |
| Maximum | 44.70 | 169.00 | 178.00 | 513.00 | 249.00 |
| Repeatability variance s_r^2 | 2.2466 | 12.1330 | 39.0164 | 76.3567 | 105.3390 |
| Intergroup variance s_r^2 | 9.61893 | 146.39249 | 151.90249 | 535.61827 | 292.14282 |
| Reproducibility variance s_r^2 | 11.8655 | 158.5254 | 190.9189 | 611.9750 | 397.4819 |
| Overall mean | 38.90 | 148.92 | 145.76 | 478.97 | 210.37 |
| Repeatability standard deviation | 1.50 | 3.48 | 6.25 | 8.74 | 10.26 |
| r Limit | 4.242 | 9.858 | 17.677 | 24.729 | 29.046 |
| Repeatability CV | 3.9 | 2.3 | 4.3 | 1.8 | 4.9 |
| Reproducibility standard deviation | 3.44 | 12.59 | 13.82 | 24.74 | 19.94 |
| R Limit | 9.748 | 35.632 | 39.103 | 70.009 | 56.422 |
| Reproducibility CV | 8.9 | 8.5 | 9.5 | 5.2 | 9.5 |
| Horwitz RSD | 6.09 | 4.97 | 4.99 | 4.17 | 4.72 |
| Horrat r | 0.6 | 0.5 | 0.9 | 0.4 | 1.0 |
| Horwitz RSD | 9.22 | 7.53 | 7.56 | 6.32 | 7.15 |
| Horrat R | 1.0 | 1.1 | 1.3 | 0.8 | 1.3 |

According to the Horrat, the repeatability and reproductibility of the method are acceptable

COMPENDIUM OF INTERNATIONAL METHODS OF WINE AND MUST ANALYSIS

Methanol (GC) (Type-IV)

| | Z-score | Z-score | Z-score | Z-score | Z-score |
|-----------------|----------------|------------------|-------------|----------------|--------------|
| Laboratory code | Dry white wine | Sweet white wine | Red wine | Oaked red wine | Port |
| A | 0.05 | -1.36 | -0.04 | 1.04 | -0.23 |
| B | 0.62 | 0.80 | 0.21 | 0.10 | 0.53 |
| C | -0.78 | -1.40 | -1.39 | -0.51 | -1.23 |
| D | -0.32 | 1.00 | 0.26 | -3.26 | 1.10 |
| E | 8.74 | 1.36 | 2.30 | 0.81 | 0.78 |
| F | -0.54 | -0.05 | -0.21 | 0.00 | -0.19 |
| G | 0.86 | 0.28 | 0.31 | 0.46 | 0.33 |
| H | -1.25 | -1.11 | -0.93 | -1.78 | -1.01 |
| I | 1.22 | 0.44 | -0.16 | -0.68 | 0.76 |
| J | 0.54 | 0.49 | 0.22 | 0.77 | 0.34 |
| K | -0.33 | -2.05 | -1.58 | -3.76 | -2.31 |
| L | -0.55 | 1.32 | 0.92 | 0.85 | 1.21 |
| M | 1.34 | 0.57 | 0.83 | 0.91 | 0.41 |
| N | -1.45 | -1.81 | -1.16 | -2.02 | -1.00 |
| O | -1.18 | 0.24 | -0.20 | -0.03 | 0.19 |
| P | 1.54 | 0.03 | -0.12 | -0.69 | -0.24 |
| Q | 0.25 | 0.20 | 0.73 | 0.76 | 0.55 |

Z-scores obtained by the participants of the 85 Z-scores, 3 are unsatisfactory and 4 are questionable